



HAL-HFX105
Handheld Formaldehyde
Meter/Monitor

Operational Manual



10302 Northridge Drive
Rancho Cucamonga, CA 91737 USA
Phone: (909) 908-3161
Fax: (909) 244-0100
Info@haltechnologies.com
<http://haltechnologies.com>



This Page intentionally left as blank

Table of Contents

Important Messages	4
I. Introduction	6
1.1 Features	7
1.2 Specifications	7
II. Basic Operation	9
2.1 Measuring Screen	9
2.2 Browsing Screen	11
2.3 Setting Screen.....	13
2.4 Calibration	14
III. Warrantee	17
Contact.....	19
User Registration Form	20

Important Messages

The information in this manual is believed to be accurate to date. However, Hal Technology assumes no responsibility for any inaccuracies that may be contained in this manual. In no event will Hal Technology be liable for direct, indirect, special, incidental, or consequential damages resulting from any defect or omission in this manual, even if advised of the possibility of such damages. In the interest of continued product development, Hal Technology reserves the right to make improvements or changes in this manual and the products it describes at any time, without notice or obligation.

Published in the United States of America

Copyright © 2008 by Hal Technology



All rights reserved. No part of the contents of this manual may be reproduced, transmitted, stored, or translated into any other language in any form by any means without the written permission of Hal Technology.

Quality Assurance

- This product has met the product specifications. All the test instruments and standard materials used for calibration are traceable.
 - This certification is for new production only and not valid for used one or ones for an exhibition purpose.
-

Commonly used symbols in this manual

Following symbols are used throughout this manual:



WARNING

The action could lead to harmful damage to the instrument.



NOTICE

Bring you attention about the features of the instrument.

Unpacking and Inspection

- Inspect the receiving package and notify the shipper immediately if there appears to be susceptible damage during shipping.

- Please verify that the enclosed items match with the shipping package list.
-



WARNING

This Instrument also contains static sensitive components that may be damaged by improper handling. The warranty is void for any unauthorized opening of the instrument.

Environmental Requirements

To avoid any accident or damage to the instrument, please avoid using in the following situations:

- DO NOT expose to combustible, explosive environments.
 - DO NOT expose to environments where rust or radioactivity are present.
 - DO NOT expose to an environment exceeding the specified limits.
-

Technical Support and Warrantee

Within a year from the date purchased, the manufacturer will provide free technical support and software upgrade if applicable. For additional help, please contact info@haltechnologies.com



NOTICE

It is strongly recommended that the instrument should be calibrated semi-annually or annually at most. Please contact Hal Technology to schedule your calibration or any services needed.

The HAL-HFX105 can only be serviced at Hal Technology or by Hal Technology's authorizing trained professionals.

I. Introduction



Formaldehyde (HCHO) is one of the most common, poisonous substances found in daily life and industry. The HAL-HFX105 handheld Formaldehyde meter is a compact personal monitor that can provide a rapid indication of hazardous airborne formaldehyde levels at the touch of a button.

The HAL-HFX105, based on reliable electrochemical sensing technology, features directly display of the formaldehyde concentration in ppm, and is easy to use. A short, low-resistance diffusion pathway with a built-in pump enables a rapid reaction time. Low power design allows long operation time. Exceptional linearity and stability allow easy and very long intervals between calibration checks. The external digital temperature and humidity sensors or pressure sensor, if applicable, allows compensation or correction for accurate measurement. A USB port provides capability of downloading stored data and possibly continuous, real-time monitoring of the environment remotely.

The HAL-HFX105 Formaldehyde Monitor is designed for use in a wide variety of applications such as furniture, floor boards, wall papers, paint, gardening, indoor decoration, construction, dye stuffs, paper manufacture, hospitals, pharmaceutical, medical, food, cleaning, synthetic resins, textile treatment, horticulture, and cosmetics.

1.1 Features

- Easy to use - minimal operator training required
- Wide measuring range
- Rapid response time
- No warm-up time
- Direct real time readings allow immediate response to results
- Reliable electrochemical sensor
- Manual and auto save capability
- Auto back light (power saving)
- High-speed USB connectivity
- Simple and easy in-field calibration
- External digital temperature and humidity sensors to assure accurate measurement
- Excess limit warning (user defined)
- No less than 6 hours of continuous operation.

1.2 Specifications




- Target Gas: Formaldehyde (HCHO) in air
- Sensor Technology: Electrochemical sensor
- Sampling Method: Pump and pointing sampling
- Range: 0 ~ 5ppm (0 ~ 20ppm available upon request)
- Response Time: < 30 seconds
- Resolution: 0.01ppm
- Long Term Draft: <10% per year
- Repeatability: < $\pm 2\%$
- Position Sensitivity: None
- Expected Sensor Life: 5 years in non-corrosive environment.
- Display Unit: ppm (4 digit LCD)
- Memory: Up to 500 sets of data
- Interface: USB




- Power: Rechargeable Lithium ion battery (3.7V/900mAh); AC adapter 100~240VAC to 5VDC/1A
- Dimension: 80 (W) × 150 (H) × 36 (D) mm
- Weight About 200 grams
- Environmental Condition: Operating: 5~ 40°C, <90%RH; Storage: -20~ 50°C, <90%RH
- Standard accessories: AC adapter, USB cable, CD with data download software and user manual
- Optional accessories: Temperature and humidity sensor probe

Sensor cross sensitivity data

Chemicals	Response (HCHO equivalent)
100ppm Methanol	1.0ppm
10 ppm Ethanol	1.0ppm
25 ppm Isopropanol	0.5ppm
20 ppm Carbon Monoxide	1.0ppm
25 ppm Phenol	0.05ppm
100 ppm Acetaldehyde	0.5ppm
100 ppm H ₂	0.5ppm
50 ppm H ₂ S	3.0ppm
20 ppm SO ₂	1.0ppm
Methane, acetone, CO ₂ H ₂ O vapor have no response	

II. Basic Operation

Six control keypads are used to operate the instrument: , **RUN/STOP**, **ENTER**, **BACK**, , .

- Power button : Push and hold it for about 2 seconds to turn on the instrument. After turned on, keep pushing on for about 2 seconds to turn off the instrument. The instrument will shut off automatically to save the power after about 8 minutes standby or no operation,
- **RUN/STOP**: Start or stop a measuring/sampling operation.
-  : Move the cursor to select desired window page or item.
- **ENTER**: Confirm the current selection or enter parameter or save current sampling value.
- **BACK**: Change the concentration unit in ppm and mg/m³ or back out of the selection

The bottom of the enclosure includes

- USB Interface: Connect to the USB interface to a computer for data downloading, remote sampling or firmware upgrading. Contact the manufacturer or sale representatives for availability of these functions.
- POWER port: An AC adapter plug-in port.
- Charge Status LED: LED flashes during charge and becomes steady after the charge finished.

2.1 Measuring Screen

The Measuring Screen is the main screen of the instrument for sampling testing. This screen can be run at the default settings. It displays the test result and conditions according to the unit settings. One may wish to change the settings before a measurement run by pressing up or down arrow to enter into the Setup Screen. An example of the Measuring Screen is shown in the Figure 1.




Figure 1 Measuring Screen

Battery Indicator

The battery indicator displays the battery strength graphically. Four bars represent 100% of charge in the battery; three bars 75%; two bars 50%; one bar 25%. No bars signify a low battery status and simultaneously the alarm will buzz as a warning. Charging of the battery is necessary at this level and after a few seconds of the warning sound the instrument shuts itself off automatically.

Time and Date

The current date and time is always displayed in the format of year-month-day and hour: minute: second, respectively. Date and time can be changed in the Setting screen.

Temperature & Relative Humidity (T and R/H)

Temperature and Relative Humidity automatically displays when an external temperature and humidity sensor probe is attached.



NOTICE

- Use **BACK** keypad to toggle the concentration unit between ppm and mg/m^3 .
- Push **RUN/STOP** keypad to start continuous measuring/sampling. During the sampling process, the backlight will be off automatically.
- Data are logged or current sampling value will be saved every time when the **ENTER** keypad is pressed. Push **RUN/STOP** keypad to stop measuring and the backlight will be on again.

- The instrument will automatically turn off the power in 8 minutes or so if no keypad action was received. However, the power will not be automatically turned off as long as a sampling (with a running pump) is in process.

Pressure

When it is applicable, the current pressure value will be displayed in Pa when an external pressure sensor is attached. The calibration data is set in the Setting screen.



WARNING

- Auto backlight is on and the instrument will automatically turn off the backlight after about several seconds if there is no keypad action.
- The instrument will be automatically turned off if there is no keypad action after about 8 minutes.

2.2 Browsing Screen

Press the arrow keypad to select the Browsing screen and enter into this page (Figure 2). This screen will allow the user to browse or delete historic data.






Figure 2 Browsing Screen

In the Browsing screen, press **ENTER** keypad to enter into the last saved data record (Figure 3). Then use the arrow keypad to scroll through the stored data.

- Use  keypad to go to the next saved data record.
- Use  keypad to return to the previous saved data record.

- Use **BACK** keypad to return to the main Browsing window. **BACK** keypad is effective only after entering into the data record.




Figure 3 An example of the Browsing Screen

Record

Record format as current number of saved data/total number of stored data. (e.g., 020/020). In the Figure 3, one views the last data in the record with the total 20 of data sets stored.

Delete the record

- Press an arrow keypad to move up or down to select the record to be deleted.
- Use the arrow keypad to select Delete one or Delete all. Then press **ENTER** keypad to delete the current record (in this case, the total number of record will be reduced one while the next record number will be moved to replace the delete one), or delete all the records.




Figure 4 An example of deleting records in the Browsing Screen

2.3 Setting Screen

The Setting screen allows users to set or change date, time, limit of warning as well as auto save mode. Use an arrow and **ENTER** keypads to enter into the Setting screen. Then press the **ENTER** keypad to highlight the parameter that needs to be changed / set. The chosen parameter will be at the bottom of the window. Press **ENTER** keypad again and using the keypads scroll through the parameter options or change the number and then press **ENTER** to confirm the parameter's setting. Press **BACK** keypad to back to the previous screen.




Figure 5 Setting Screen

Alarm level setting

The user may turn on or off the excess exposure limit warning. The user may input any value between 0 and 9.99ppm. For your reference, two levels of limit for Formaldehyde are in common: 1) Threshold Limit Value (TLV) of 0.3 ppm as a "ceiling limit" established by American Conference of Governmental Industrial Hygienists (ACGIH); 2) Short Term Exposure Limit (STEL) of 2 ppm (15 minute average) established by US Occupational Safety & Health Administration (OSHA). The Immediately Dangerous to Life and Health (IDLH) limit is 19ppm.

Save

Setting the parameter for **SAVE** can turn on or off the automatic saving function. The parameter **M** represents manual save mode (auto save mode off). In this condition, the reading will be logged or a current sampling value can be saved every time when the **ENTER** keypad is pressed. If the setting is the number n (n is 1 to 9), the instrument will automatically save the sampling value every n minute. For example, n

=5, it means that the instrument will save a current value every 5 minutes.



NOTICE

- While the automatically saving function is turning on, one can still press the **ENTER** keypad to save the current sampling value.

2.4 Calibration

After turning on the instrument, use up or down keypad to go to the Calibration Screen (Figure 6). Users may calibrate the instrument at his/her own wish after using for a certain time of period or suspect degradation of sensor performance.

Calibration			
[X]	[Y]	[K	B]
0.00	0000	+	208
1.00	0610	+	000

Figure 6 Calibration Screen -1





Recommended calibration method is standard Zero-Span technique. The X values in the first column represent the concentration of calibrating gas in ppm. The Y values in the second column represent the response of the sensor to be calibrated. The values in the third column are calibration coefficients. If the 5ppm sensor is used, default values of calibration coefficients, K and B, are 208 and 0, respectively. The example of calibration procedures is described below:

Here we use a standard zero-air gas cylinder and the vapor concentrations of 1 ppm of formaldehyde standard solution as an example for calibration:

- 1) Prepare a glass of 1ppm standard formaldehyde (HCHO) solutions and fill the beaker no more than one third of its volume.
- 2) Connect the inlet of the instrument to a glass container and introduce the zero-air gas to the glass container.
- 3) Enter into the calibration screen by pressing the **ENTER** keypad and make sure that the left arrow pointed at the first row of 0ppm concentration. Use up or down key to move the cursor (refer to Figure 7).

Calibration			
[X]	[Y]	[K	B]
0.00	0000	+	208
1.00	0610	+	000
Reset		0.00 ppm	

Figure 7 Calibration Screen -2

- 4) Press **RUN/STOP** keypad to start a sampling. Wait for right bottom number stable and then press **STOP** key.
- 5) Connect the inlet to a vapor out of the standard liquid (e.g., 1ppm standard formaldehyde liquid solution).
- 6) Move the cursor to the second row (refer to Figure 8) and then press the **ENTER** keypad. Use  or  to select the element to be changed. When the element of the X column becomes highlighted, use  or  to change the number and press **ENTER** to confirm the change. Set the number as the concentration level of the standard liquid to be tested (e.g., 1.00ppm). Then use **BACK** keypad to return to the main Calibration screen.

Calibration			
[X]	[Y]	[K	B]
0.00	0000	+	152
1.00	0310	+	067
Reset		0.00 ppm	

Figure 8 Calibration screen- 3

- 7) Press **RUN/STOP** keypad to sample, the 2nd-row value in the Y column will display the response of the concentration for current liquid standard. Press **RUN/STOP** keypad again to stop the sampling after the reading is stabilized.
- 8) After finished both samplings on standard gases, move the cursor back to the first row, press and hold the **ENTER** key for about two seconds. The instrument will automatically calculate and update the calibration coefficients based on new settings and calibrations. Press **ENTER** again to save values.
- 9) After finishing calibration procedures, press **BACK** to exit.



NOTICE

To restore the default settings of factory calibration, move cursor to highlight **RESET** and then press **ENTER** to restore factory calibration for coefficients of K and B. press **BACK** key to exit.

III. Warrantee

Hal Technology provides a one-year limited warranty of the Model HFX105 Handheld formaldehyde meter, but not including necessary calibration service.

- Warranty begins from shipping date.
- The user is responsible for the cost of shipping in the case of any service or repair needed.
- The warrantee only limits to the HFX105 and HAL TECHNOLOGY does not extend this liability to accessories and any other equipment damage, body injury and loss of properties due to abnormal use.

The following are not included in the warranty:

- Improper connection to a power source, resulting in damage of the instrument.
- Any physical damage due to mechanical forces (e.g., collision or dropping) that may cause any damage of the front panel, LCD screen, switch and internal components, etc.
- Unauthorized opening of the instrument.
- Damage due to operation in an un-specified environmental condition.
- Abnormal operation due to instrument needing calibration.

Limitation of Warranty

A. Hal Technology warrants that all equipment shall be free from defects in material and workmanship under normal use for a period of one year from date of shipment to Buyer except that Hal Technology does not warrant that operation of the software will be completely uninterrupted or error free or that all program errors will be corrected. Buyer shall be responsible for determining that the equipment is suitable for Buyer's use and that such use complies with any applicable local, state, or federal law. Provided that Buyer notifies Hal Technology in writing of any claimed defect in the equipment immediately upon discovery and any such equipment is returned to the original shipping point, transportation charges prepaid, within one year from date of shipment to Buyer and upon examination Hal Technology determines to its satisfaction that such equipment is defective in material or workmanship, i.e. contains a defect arising out of the manufacture of the equipment and not a defect caused by other circumstances,



including, but not limited to accident, misuse, unforeseeable use, neglect, alteration, improper installation, improper adjustment, improper repair, or improper testing, Hal Technology shall, at its option, repair or replace the equipment, shipment to Buyer prepaid. Hal Technology shall have reasonable time to make such repairs or to replace such equipment. Any repair or replacement of equipment shall not extend the period of warranty. If the Instrument is modified or in any way altered without the explicit written consent of Hal Technology then the warranty is null and void. This warranty is limited to a period of one year, except as noted below, without regard to whether any claimed defects were discoverable or latent on the date of shipment.

B. If Buyer shall fail to pay when due any portion of the purchase price or any other payment required from Buyer to Hal Technology under this contract or otherwise, all warranties and remedies granted under this Section may, at Hal Technology's option, be terminated.

C. Warranty repairs shall be completed at a Hal Technology authorized service location, by an authorized service technician, or on site at buyer's facility by a Hal Technology authorized employee. Buyer pays shipping costs to factory; seller will pay standard return shipping costs during the warranty period. Buyer may select a faster method of shipment at their own expense.

Warranty of Repairs after Initial One (1) Year Warranty

A. Upon expiration of the initial one-year warranty, all parts and repairs completed by an authorized Hal Technology repair technician are subject to a six (6) month warranty.

B. Other than the above, Hal Technology makes no warranty of any kind, expressed or implied, except that the products manufactured and sold by Hal Technology shall be free from defects in materials and workmanship and shall conform to Hal Technology's specifications; Buyer assumes all risk and liability resulting from use of the products whether used singly or in combination with other products. If instrument is modified or in any way altered without the explicit written consent of Hal Technology, then the warranty is null and void.

C. Warranty repairs shall be completed at a Hal Technology authorized service location, by an authorized service technician, or on site at buyer's facility by a Hal



Technology authorized employee. Buyer pays shipping costs to factory; seller will pay standard return shipping costs during the warranty period. Buyers may select a faster method of shipment at their own expense.

Contact

HAL TECHNOLOGY, LLC
10302 Northridge Drive
Rancho Cucamonga, CA 91737 USA
Phone: (909) 908-3161
Fax: (909) 244-0100
Info@haltechnologies.com
<http://haltechnologies.com>

Information Record	
Model	_____
Serial No.	_____
Purchase Place	_____
Address	_____ _____
Phone	_____
Service Place	_____
Address	_____ _____
Phone	_____



Please fill out the Registration form below and send to:

HAL TECHNOLOGY, LLC
10302 Northridge Drive
Rancho Cucamonga, CA 91737 USA
Phone: (909) 908-3161

Or send relevant registration information to the email address below:

info@haltechnologies.com

<p>User Registration Form</p> <p>Company _____</p> <p>Contact Person _____</p> <p>Address _____</p> <p>_____</p> <p>City _____ State/Province _____ Country _____</p> <p>Postal Code _____</p> <p>Phone _____</p> <p>Fax _____</p> <p>E-mail _____</p> <p>Product Model _____</p> <p>Serial No. _____</p> <p>Purchase Date _____</p> <p>Purchase Place _____</p> <p>Preferred Contact Method</p> <p><input type="checkbox"/> E-mail <input type="checkbox"/> Mail <input type="checkbox"/> Phone</p>
